


SN# 001517 Lamp Specific Acceptance Test



Model: UV-3000, Magnum-GO	Result	Steady State	Acceptable	Units
Peak Output Wavelength	Pass	368	365 ± 5	nm
Full Width Half Maximum (FWHM)	Pass	10	≤ 15	nm
Full Width at 10% Maximum	Pass	21	≤ 30	nm
Longest Wavelength at Half Maximum	Pass	374	≤ 377	nm
Max Excitation Irradiance (347-382 nm)	Pass	3657	≥ 2000	$\mu\text{W}/\text{cm}^2$



Inspector: REL
 Date of Certification: 6-16-2026

This lamp has been tested to conform to all requirements of:
 ASTM E3022 ✓ AITM6-1001 ✓ ISO 3059 ✓ RRES 90061 ✓

Emission Spectrum Measured at Stabilization Temperature with Blue-wave UV-25 S/N 18120729, Certificate Number 18120729-UV-CR2-0925



C4 Magnum Series, Certificate of Conformance



SN# 001517	Switch On	Steady State	Elevated Temp*	Acceptable	Units
Peak Output Wavelength	367	368	370	365 ± 5	nm
Full Width Half Maximum (FWHM)	10	10	13	≤ 15	nm
Full Width at 10% Maximum	21	21	27	≤ 30	nm
Longest Wavelength at Half Maximum	373	374	376	≤ 377	nm
Visible Emission at 15 in	9	8	8	≤ 20	Lux
Visible Emission at 36 in	3	3	3	≤ 5	Lux
Irradiated Area at 15 in ($>1000 \mu\text{W}/\text{cm}^2$)	11	10	9	≥ 5	Inch Diameter
Max Excitation Irradiance (347-382 nm)	3,840	3,657	2,400	$\geq 2,000$	$\mu\text{W}/\text{cm}^2$
Irradiance at Peak Output Wavelength	340	324	185	-	$\mu\text{W}/\text{cm}^2$
Min Working Distance** ($< 5000 \mu\text{W}/\text{cm}^2$)	14	12	11	-	Inches
Max Working Distance** ($> 1200 \mu\text{W}/\text{cm}^2$)	26	23	21	-	Inches
Body Stabilization Temp (70°F Ambient)	-	116	-	≤ 120	°F
Current Ripple	9	9	9	≤ 35	mA peak-peak

* Tested in a 110°F Environment, the maximum ambient temperature where light still meets all requirements

** C4 Magnum series lights use a single LED emitter, thus beam uniformity does not change with working distance (RRES 90061)

